

Temperature Log for Refrigerator and Freezer — Fahrenheit

Month/Year: _____ Days 1–15

Completing this temperature log: Check the temperatures in both the freezer and the refrigerator compartments of your vaccine storage units at least twice each working day. Place an “X” in the box that corresponds with the temperature and record the ambient (room) temperature, the time of the temperature readings, and your initials. Once the month has ended, save each month’s completed form for 3 years, unless state or local jurisdictions require a longer time period.

If recorded temperature is in the shaded zone take immediate corrective action:

This represents an unacceptable temperature range. Follow these steps:

1. **Move vaccine(s) to a working storage unit.**
2. / ~~DEHOMHYDFHGHV DV~~ “do not use”, do NOT destroy/discard the vaccine(s).
 . \$ ~~FMYDM\RXUIDFLOMHYDFHGH(P HJ HQF 5 HMIHYDODQG6WRUJH3DQ~~
 . **Contact *USAMMA/DOC as well as your Immunization Healthcare Specialist (IHS)** and standby for further instructions on the disposition of the vaccine.
5. **Document the action taken** on the reverse side of this log.

Day of Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	
Staff Initials																
Room Temp.																
Exact Time																
°F Temp	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm
Refrigerator temperature	≥49°	Take immediate corrective action if temperature is in shaded section*														
	48°	Take immediate corrective action if temperature is in shaded section*														
	47°	Take immediate corrective action if temperature is in shaded section*														
	46°															
	45°															
	44°															
	43°															
	42°															
	41°															
	40°															
Freezer temp	39°															
	38°															
	37°															
	36°															
	35°	Take immediate corrective action if temperature is in shaded section*														
	34°	Take immediate corrective action if temperature is in shaded section*														
Freezer temp	33°	Take immediate corrective action if temperature is in shaded section*														
	32°	Take immediate corrective action if temperature is in shaded section*														
	31°															
	30°															
	29°															

Aim for 40°

*USAMMA/DOC Emergency Contact: Phone: (301) 619-4318, DSN (343), After hours: (301) 676- , Hmail: ~~XMDP\GHEMNP HGRP XMDP P DP EJ GR# P DCP IO~~

Vaccine Storage Troubleshooting Record

Use this page to record the details of the vaccine storage incident, including the date and time of the last known temperature within the appropriate vaccine storage range.

Date	Time	Storage Unit Temp	Room Temp	Incident	Action Taken	Results	Initials

Temperature Log for Refrigerator and Freezer — Fahrenheit

Month/Year: _____ Days 16–31

Completing this temperature log: Check the temperatures in both the freezer and the refrigerator compartments of your vaccine storage units at least twice each working day. Place an “X” in the box that corresponds with the temperature and record the ambient (room) temperature, the time of the temperature readings, and your initials. Once the month has ended, save each month’s completed form for 3 years, unless state or local jurisdictions require a longer time period.

This represents an unacceptable temperature range. Follow these steps:

1. **Move vaccine(s) to a working storage unit.**
2. **Label the vaccine(s) as “do not use”, do NOT destroy/discard the vaccine(s).**
3. Activate your facilities vaccine Emergency Retrieval and Storage Plan.
4. **Contact *USAMMA/DOC as well as your Immunization Healthcare Specialist (IHS)** standby for further instructions on the disposition of the vaccine.
5. **Document the action taken** on the reverse side of this log.

If recorded temperature is in the shaded zone take immediate corrective action:

Day of Month	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
Staff Initials																	
Room Temp.																	
Exact Time																	
°F Temp	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	am pm	
Too warm*	≥49°																
	48°	Take immediate corrective action if temperature is in shaded section*															
	47°																
	46°																
	45°																
	44°																
	43°																
	42°																
	41°																
	40°																
Too cold*	39°																
	38°																
	37°																
	36°																
	35°																
	34°																
	33°	Take immediate corrective action if temperature is in shaded section*															
	32°																
	Too warm*	8°															
		7°	Take immediate corrective action if temperature is in shaded section*														
6°																	
5°																	
4°																	
3°																	



Refrigerator temperature

*USAMMA/DOC Emergency Contact: Phone: (301) 619- 4318, DSN (343), After hours: (301) 676- , Hmail: XVDU \ GHM\ENP HFRP XVDU P DP E\ GR# P DCP IO

Vaccine Storage Troubleshooting Record

Use this page to record the details of the vaccine storage incident, including the date and time of the last known temperature within the appropriate vaccine storage range.

Date	Time	Storage Unit Temp	Room Temp	Incident	Action Taken	Results	Initials